



*EvaSoft Dental Imaging Software
Users Manual*



<i>Revision</i>	<i>Effective Date</i>	<i>Description</i>
1.0	September 1, 2010	Initial Draft
2.0,2.1,2.2	Not Applicable	These revisions not released.
2.3	February 15, 2012	Update to Version 2.3 Added description of feature and GUI changes Added regulatory information
3.0	October 15, 2012	Update to Version 3.0 Revised feature and GUI changes Revised regulatory information
3.2	May 1, 2013	Eliminated CE reference Included EVA Select
3.3	May 1, 2014	Updated Company name Updated GUI graphics Added EVA, EVA Select Supplemental Information Remove Durability Section
3.3.1	August 1, 2014	Update Address
3.4	March 31, 2015	Update to version 3.4 software Update Email, Explain Study/History Modes Eliminated EVA Classic Supplementary Information Appendix
3.5	December 31, 2016	Improved Patient Tab Select and Display Behavior Scheduled Automatic Data Back Up Utility
4.0	July 1,2018	Simplified Interface Enhanced DICOM functionality

Copyright notice

Copyright©2018 SMK Imaging LLC dba: ImageWorks. All Rights Reserved. This manual shall not, wholly or in part, in any form or by any means be reproduced or transmitted without prior written authorization from ImageWorks.

This manual has been produced to assist in providing instruction for EVAsoft and EVA digital sensor products. Every effort has been made to make the information in this guide as accurate as possible. The authors of this guide shall have neither liability nor responsibility to any person or entity with respect to any loss or damages in connection with or arising from the information contained in this manual.

Trademarks

EVA, Dent-X, and ImageWorks are registered trademarks of SMK Imaging LLC. Any additional software products named in this document are claimed as trademarks or registered trademarks of their respective companies.

Regulatory Notes

ImageWorks provides you with the following information regarding the regulatory status of EVAsoft, EVA, and EVA Select:

United States of America

EVAsoft Digital Imaging Software (the application software) is included in the premarket notification 510(k) for the EVA and EVA Select products which were submitted to the United States Food and Drug Administration and received clearance as Class II Medical Devices. The K number for the EVA product is K030647. The K number for the EVA Select product is K123419.

Canada

EVA, EVA Select and the application software (EVAsoft) as EVA Digital Dental X-Ray System is licensed for sale and use in Canada. This product possesses a Class 2 Medical Device License issued 12-31-2010, by the Therapeutic Products Programme of Health Canada.

Responsibility for equipment

management attributable to the user:

Each state, province and country may have specific requirements to operate the equipment. The intended users of this equipment should be medical professionals who are familiar with and comply with the relevant local laws and regulations when operating the equipment. The user is also required to use and maintain this equipment as described in this manual.

Computer Requirements:

The computer requirements are documented separately. Please contact your dealer, ImageWorks, or our website at www.ImageWorksCorporation.com for a copy of these requirements.

Required Maintenance and Inspection:

EVAsoft Software is designed to be operated on a PC type computer running one of the approved operating system versions.

<i>Daily:</i>	Backup data (database and images at a minimum)
<i>Monthly:</i>	Upload and apply Operating system patches and security fixes. Check the free capacity of the hard drives particularly on the server/host. Check the validity of the daily backups.
<i>Annually:</i>	Check for updates on the ImageWorksCorporation.com web site. Review the list of users and, if necessary, passwords.

Actions To Be Taken In The Case Of An Accident:

1. In the event of an accident immediately stop using the equipment.
2. Provide appropriate first aid and other emergency medical assistance as required by the accident conditions.
3. If necessary, turn off and disconnect the equipment. Shutdown and power down the computer.
4. Immediately contact the dealer where the equipment was purchased. If you cannot contact the dealer, contact ImageWorks.

TABLE OF CONTENTS

REVISION RECORD and REGULATORY NOTES	2
COMPUTER REQUIREMENTS	3
REQUIRED MAINTENANCE	3 IN
CASE OF ACCIDENT	3
TABLE OF CONTENTS	4
ABOUT THIS PRODUCT	5
Product Registration	5
EVASOFT OVERVIEW	5
EVASOFT LOG-IN FEATURE	6
DETAILED TAB DESCRIPTIONS	8
<i>Patient Tab</i>	8
Stack and Work Areas	8
Create a New Patient	9
Patient/Image Searching	10
Selecting the Active Patient and Study	11
Patient Tab Toolbar	11
<i>Acquire Tab</i>	12
Capturing an Image	12
Acquire Tab Toolbar	14
<i>Diagnosis View Tab</i>	15
Annotation Tools/Demographic Overlays	16
Diagnosis View Toolbars	17
<i>View All Tools and Tooth Number Editor Windows</i>	19
Chart View Tabs and Toolbar	20
THE IMAGE TRANSFER SYSTEM	21
EVASOFT FILE MENU	23
EVASOFT EDIT MENU	23
EVASOFT IMAGE DEVICES MENU	23
EVASOFT TOOLS MENU	24
DICOM Configuration Settings Menu	25
EVASOFT PREVIEW and HELP MENU	26
ENHANCEMENT TOOL DESCRIPTIONS	27
Angle Measurement	27
Brightness/Contrast/Gamma	27
Crop	27
Despeckle Filter	27
Edge Enhance	27
Equalization	27
Flip/Rotate	27
Gamma	27
Invert	27
Linear Stretch	27
Local Linear Stretch	27
Image Calibration	27
Image Measurement	27
Sharpen	28
Soften	28
USM	28
Window/Level/Gamma	28
Pseudo Color, Histogram and Threshold	28

APPENDIX A: Layout Editor	30
APPENDIX B: Print PDF Reports	31
APPENDIX C: Network Setup	33
APPENDIX D: Tools> options	35
APPENDIX E: EVA SELECT and EVAsoft	46

ABOUT THIS PRODUCT

THEORY OF PERFORMANCE

The EVAsoft Dental Imaging Software application is designed to acquire, display, organize, store and output digital radiographic images for dental diagnostics. The software is a useful interface for all ImageWorks' digital dental X-Ray imaging equipment. In order to capture radiographic and other images, EVAsoft must have an active patient selected. Once the digital radiographic image and/or images are acquired, EVAsoft will store the images in the patient database as DICOM files. After an image has been acquired it is automatically saved and displayed on a computer monitor. A suite of image processing tools are available for the operator to enhance the displayed image for their particular diagnostic analysis. Such enhancement tools include, but are not limited to: Magnify, Gamma, Window and Level, Linear Stretch, and Equalize. Each anatomical template used to acquire digital X-ray images has a layout for viewing the images together. The Chart View tab allows the user to view the selected patient's images in these pre-defined layouts. EVAsoft also implements specific DICOM operations. Please refer to the EVAsoft DICOM Conformance Statement for details.

IMPORTANT CONSIDERATIONS:

The software is intended for dental professionals, and not intended to replace the skill and judgment of a qualified dental practitioner. The software should only be used by people who have been appropriately trained in the software's functions, capabilities and limitations. The users should be aware of the limitations in the accuracy and correctness of the output data displayed on the screen, printed, or exported from EVAsoft. The quality of the image and text data is dependent on quality and accuracy of the input data, the user's interaction with the data, the quality, characteristics, and settings of the display device or printer, and the necessity to interpolate the data for display purposes.

While EVAsoft has been extensively tested, it is impossible to completely test any piece of software, and errors may remain in the software. It is possible that an error could manifest as an incorrect measurement or image. Users must be aware of the potential for errors. EVAsoft saves images together with patient information (as DICOM files), both when saving to the local database and when exporting from EVAsoft. It is important and in some jurisdictions legally required that this data is protected from access by unauthorized persons.

Users should be aware that certain views make use of interpolated data. This is data that is created by EVAsoft based on the original data set. Interpolated data may give the appearance of healthy tissue in situations where pathology that is near or smaller than the imaging/display resolution may be present. On occasion, interpolated data may also include image artifacts which should not be interpreted as real pathology.

PRODUCT REGISTRATION

EVAsoft requires a license to use beyond the 30 day full version trial period. To register your EVAsoft Software via the internet, go to Help > License Registration, or call ImageWorks' technical support and register over the phone. Note if you are using EVAsoft in networked mode only the Host requires registration.

CONTACT INFORMATION

Technical Support: 1 (914) 592-6100
E-mail: support@imageworkscorporation.com
Web: www.ImageWorksCorporation.com

EVASOFT OVERVIEW

EVAsoft has been designed using a tabbed interface. Each tab is designated a specific set of tasks. There are two general areas within each tabbed area. To the left is the main imaging area (or screen). To the right is an expandable and collapsible smaller area referred to as the “Stack”. The “Stack” usually contains certain tools or options that directly pertain to the purpose of the main tab that it falls under. For example, patient search criteria are found under the Stack of the PATIENT tab. Tooth charts, patient information, exposure details, etc. are found under the Stack of the ACQUIRE tab. Tools, demographics and annotation tools are found under the Stack of the DIAGNOSIS tab. There are also sub-tabs present in both the main window and sub windows as well as in the Stack areas.

The **Patient Tab** is designed for designating the active patients, as well as accompanying active studies. Additions of new patients (either with full information or emergency patients whereby their information may be filled in later), editing existing patient information and conducting various searches under a wide variety of criteria are all possible within this section.

The **Capture Image Tab** is designed for the purpose of acquiring images. The user is presented with a number of acquisition templates from which to choose from a drop-down list in the Stack. From the template, the user may decide to remain with the default sequence of exposures or select a new one. In


addition to sequential mode is the Endo mode. This allows the user to take exposures on a given tooth one at a time or select the sequence on-the-fly.

The **Exam View Tab** is designed to view images from a study in a predefined layout. The user can also drag and drop images into different positions on the chart.


The **Diagnosis View** is designed to allow the user to view and edit one image at a time. All of the image enhancement tools and filters are located in this tab, as well as annotation tools and demographic data overlays. This view is enabled by double clicking on any image.

EVASOFT LOG-IN FEATURE


EVAsoft has the optional use of login accounts. It is recommended that each user create an account name and password. At start-up, EVAsoft displays the Login window shown below. The login system can be set up to require user names and passwords, user names without passwords, or no user name or password.



To add user accounts and/or edit user accounts, including setting up passwords, go to **Edit > User Accounts**



To make changes to the User Login Preferences, go to **Tools > Options**, in the General preferences area.



For further explanations of the Options Window, see **Appendix D**.

DETAILED TAB DESCRIPTION


STACK AND WORK AREAS

Each tab has two main working areas. On the right-hand side of the area is what is called the “Stack”. The Stack may be collapsed and expanded as needed and has various subjects that are applicable to whatever tab is selected. For example, in the Patient tab, the Stack on the right has search criteria for sorting through patients, ID numbers, etc. One of the features of this Stack is the ability to display DICOM Worklist Patients.

The area on the left is the working area. Both the working area and that Stack area have their own smaller tabs. The Diagnosis View tab, in particular, has a number of smaller tabs in the Stack area that pertain to tools, demographics, annotations, etc.


HOME TAB

The Home Tab is the entry point into the software. It also serves as a “home base” to return to at any point in the workflow.



SELECT PATIENT TAB

In this tab the user can create new patients, find existing patients, and edit patient information. When the program is launched, the default view is “Recent Patients” which can be found among the buttons that are on the right-hand side. This will display only the patients that were accessed the last time the program was run. Clicking on “Show All Patients” in the right-hand Stack will bring up all existing patients.





To create a new Patient, click the New Patient button. The following window will open where the user can enter all the patient information. Make sure to hit the “enter” key after entering the patient’s birth date.

Enter new patient information ...

Please enter a unique Alphanumeric ID. Special Characters are allowed.

* Patient ID: 61378098410712

Flag: ☐

* First Name:

Middle Name:

Last Name:

Gender: ☒ Male ☐ Female

Date of Birth: 02/14/2012

Social Security Number:

Dentist: Einstein, Albert

Referring Physician Name:

Day Phone:

Evening Phone:

Mobile Phone:

Address 1:

Address 2:

City:

State / Province:

Zip / Postal Code:

Country:

E-mail:

Insurance:

Comments:

Create Patient Cancel



Patient/Advanced Search Search tools are divided into two categories: quick search buttons and detailed searching. The following quick searches are available:

Quick Search

DICOM Worklist Patients

Deleted Patients Only

Flagged Patients

Anonymous Patients Only

Patients Over Last Month



Recent Patients

Show All Patients



The user can search for the Patient Name and/or Patient ID by entering the text in the corresponding text box and clicking the “Find” button.

- Only patients with the matching criteria will be displayed. The user can also choose dates that images were captured on or in between, again clicking the “Find” button to apply the search. Only patients with images captured on the selected date or in the selected date range will be displayed.
- Use the Dentist, Tooth Number and Image Comment searches to only display patients that match those criteria.
- Once a search is complete, the list of patients will be displayed. Click on the Plus (+) sign next the patient’s name to see that patient’s list of studies and images.

  Monroe, Marilyn

Selecting the Active Patient and Study

- Start in the “Search Results” sub-tab.
- To select the Active Patient, the user must click on the patient’s name. To select the Active Study, the user must single-click on the study name.
- To launch the study in the View Exam Tab the user must double-click on the study.
- Right-clicking on an active study will bring up the “Edit Study Name” button and display the DICOM Study accession number (for PACs Worklist Integrations).
- Left-clicking on that button will bring up the “Edit Study Name” window.
- Use the Clear button in Advanced Search to deselect the Active patient and study.



Click on the Patient Information button to edit or delete that information.

Coco Chanel

Patient ID: 512985698745

Number of studies: 2

Last study on: Wednesday, February 1, 2012 3:19 PM

2nd to last study on: Wednesday, February 1, 2012 3:19 PM

Dentist: Einstein, Albert

Gender: F

Social Security Number:

Date of Birth: May 3, 1978

Referring Physician Name:Dr. Roberts

Comments:

Address:
987 Fancy Lane, New York, NY 55555

Phone (Day): 555-555-5555

Phone (Evening):

Phone (Mobile):

E-mail:cchanel@imageworkscorporation.com

Insurance:


Edit Patient

Delete Patient

Merge Studies

Transfer Study

Delete Study



The “Merge Study” button allows one to combine multiple studies into a single study.


The “Transfer Study” button allows one to transfer a study from one patient to another.

Click once on a study and the Thumbnail images will be displayed in the Filmstrip and the patient associated with that study will become the Active Patient as well.

CAPTURE IMAGE TAB

The Capture Image Tab is where the user can acquire images for the active patient. If no active patient is selected, the Capture Module will prompt the user to select a patient or create an anonymous patient to continue.

In order to acquire images, there must be a selected hardware device. If a device has already been selected, it will be displayed in the title bar of the EVAsoft window. If no device has been selected, the title bar will display “View Station”. See below for more details on selecting a device.



Capturing an Image

The user must complete four steps in order to capture a digital X-Ray image. The following steps will demonstrate using an EVA sensor.

After opening the EVAsoft software and logging-in, the user must:


1. Select the active patient in the Patient Tab.
2. Select the correct hardware from the Image Devices drop-down menu.
3. Go to the Acquire Tab area and select an acquisition template from the menu in the Stack (if something other than the default tooth template is desired).

4. Select the acquisition mode: Either single shot  or sequential  (FMX).

The first step: Selecting a patient. In the Patient Tab, find the desired patient and double-click on the patient's name. This patient is now

set as the Active Patient. (See page section on "Selecting the Active Patient and Study" above for detailed description of setting the active patient).

The second step: Select the hardware device that you will be using. This step is not often required because the software remembers the last device that was selected. *Go to the Imaging Devices drop-down menu:*



Select the desired hardware, for example, "EVA". The active hardware will be displayed in the title bar of EVAsoft.

Note: Only hardware that is available on the system will be displayed in this window.

The third step: Select an acquisition template. This must be done from within the Acquire Tab. The user can choose an Acquisition Template from the drop down list. Once a template has been selected, the user can remain with the default sequence of exposures or change the sequence. The default template is determined based on the age of the active patient. If a patient 12 years old or younger is active, the pedo template is shown.



Important Note: *When a template is displayed, the first region to be taken will have the numeral "1" shown on it. This will be the first of a predetermined order of exposures that are determined in the "Layout Editor" (which will be discussed later). This order may be altered immediately prior to taking exposures by clicking on the regions desired. Then, upon entering the Sequential mode, the X-rays will be taken in that new order.*

In Single Shot mode, subsequent exposures will be taken on the same tooth position until/unless a different position is chosen prior to the next exposure.

The fourth step: in acquiring an image is to select an acquisition mode: Single Shot or Sequential.








If the user wants to change this, simply click on the "Stop Acquisition" button in the Acquire Tab and reselect the mode desired.

Once an Acquire button is pressed, the software will tell the hardware to be ready to acquire an image, and the Acquire Status will change to green when ready. Now the user can operate the X-Ray.

Once the image data is received it is displayed and a thumbnail of the image is added to the filmstrip.


Each acquire session will be saved as a new study for the Active Patient. The exception to this rule is if the Active Patient has a study that is not more than 2 hours old. In this case, the user will be asked, upon entering the Acquire Tab, to append to the last study or start a new one.

ACQUIRE TAB TOOLBAR

	<i>"Start Single Acquisition Mode" Button</i>
	<i>"Start Continuous Acquisition Mode" Button</i>
	<i>"Stop Acquisition" Button</i>
	<i>"Retake Exposure" Button</i>
	<i>"Skip Ahead to Next Exposure" Button</i>



Status window



Acquisition Template



Reset Template

DIAGNOSIS TOOLBAR

At any time the user can double-click on a thumbnail from any filmstrip and it will automatically open in the Diagnosis View Tab in Study Mode. Thumbnails for all the images in the currently selected study are displayed in the filmstrip along the bottom of the screen. To view images of the same tooth position from different studies (for example from this year and last year) double click on an image position in the Chart Tab and the image will open in Diagnosis View Tab in History Mode. Study Mode has a **Gray** filmstrip and History Mode has a **Blue** filmstrip. In History mode, x-rays from all dates taken in the selected chart position are displayed in the filmstrip.

Images can be loaded into the viewing panel(s) by clicking on the panel until a green border is displayed, then clicking on the tooth image from the film strip.



Above the image are several tool buttons. To the right of the image is the Stack. The Stack contains all the image manipulation tools, annotation tools and demographic information overlay tools. The Stack area is minimized by default to allow the image to have maximum viewing area.



Simply click on the collapse/expand buttons.



To apply an enhancement tool from the Toolbox, click on the add button and select the tool from the list that appears. This will add the tool to the Stack. The Stack performs each tool in order, from top to bottom. Full descriptions of each tool can be found in the Enhancement Tool Descriptions section of this document.

Drop-down list of tools that appears in the Stack from clicking on the “Add a new operator to the Stack” button

Angle Measurement
Brightness/Contrast/Gamma
Crop
Salt & Pepper Filter
Edge Enhance
Equalization
Flip/Rotate
Gamma
Invert
Linear Stretch
Local Linear Stretch
Physical Image Calibration
Physical Image Measurement
Sharpen
Soften
USM
Window/Level/Gamma



Enable all operators in the Stack




Disable all operators in the Stack



Remove all operators from the Stack

Here is an example of the Gamma tool being chosen. Note that it is highlighted with a red bar that indicates it is active.



After each tool has been adjusted, the user can click on the red bar and it will become gray and inactive. The user can click on the “Enable All” or “Disable All” buttons which will toggle all enhancements on and off.

ANNOTATION TOOLS TAB

To apply annotation tools, click on the Annotations Tab in the Stack area. Then click the Add button and you will get a list of Annotation tools to choose from.

Add Button

Applied Tools

Annotations

Demographic Info

Annotations Tab

Demographic Info Tab

Pointer Note

Text Note

Box

Line

Left Marker

Right Marker

Oval

Annotations List

DEMOGRAPHIC DATA OVERLAYS

To apply demographic data overlays onto the image, click on the Demographic Info Tab in the Stack area. Then click on the Add button and select the “Demographic Overlay” menu item. The demographic overlay tool will be added to the Stack. The user can choose what data information to display from the Demographic Overlay Window drop-down list. The user can also choose what corner of the image to display the information. Click on “Done” to finish the operation.

After the desired operators have been applied to the image, click the save button to save the changes. This save action will add the tools from the Stack into the patient database. The next time the user opens that image, the tools will automatically be re-applied to the image, and added to the Stack.

Important Note: No changes are ever made to the actual Image File! All tools that are applied to the image are saved in the database. When the image is opened again, the tools are re-applied to the original image data.

To see the original image again, the user can simply turn off all the items in the Stack by clicking on either the “Disable all Stack items” button or... “Modified/Original” button.

GROUPING ENHANCEMENT TOOLS FOR FUTURE RE-USE

The user can save the enhancement tools and the adjustments made to those tools so that they can be applied to other images. To save the grouped set of the enhancement tools which are currently in the Stack, simply right-click on the image and select the “Send to...” option from the pop-up menu. The “Transfer Image...” window will open. Click on the “Custom Enhancement” button. “Add New” will appear. The “Custom Enhancement Label” window will open where the user can assign a name to the new group of enhancements. Then click OK.













When a new image is opened in the Diagnosis View Tab, the user can click on the “Apply Custom Enhancements” button (under the All Tools wrench icon button) and select the name of the group to apply to the image. The “Apply Custom Enhancements” button will remember the last group that was used and show that one at the top of the list until a different

19

group is selected.

This grouping also applies to Demographic Information Overlays. Add the Demographic information to the Stack and then right-click on the image. In the “Transfer Image...” window, choose the “Custom Overlay” button. Enter a name and click OK.

DIAGNOSIS VIEW TOOLBARS (UPPER)

Save Current Image	
Print the Current Image	
View Modified/ Original Image	
Undo/Redo	
Zoom Fit	
Draw Zoom	
Magnification Spotlight	
Caries Spotlight	
Increase Gamma	
Decrease Gamma	
Brightness/ Contrast*	
View All Tools	

*Clicking on Brightness and Contrast allows to hover mouse over image and make adjustments by clicking and dragging mouse. Click and drag vertically to adjust brightness. Click and drag horizontally to adjust contrast.

[VIEW ALL TOOLS WINDOW](#)


The View All Tools Window features a comprehensive set of tools. Two unique tools are the “Toggle Annotations On/Off” and the “Edit Tooth Numbers” tools.



Edit Tooth Numbers

Toggle Annotations On/Off

Clicking on the Edit Tooth Numbers button brings up the following window:



VIEW EXAM TAB

Once a study of images is captured, they can be viewed in the View ExamTab. If the images were acquired using one of the Acquisition Templates, then they can be viewed in a predefined layout that has been designed for that template. This will allow the user to see the images as if they were in an image chart. The user can also select any study for the active patient from the study drop-down list located in the upper tool bar.








CHART VIEW TAB TOOLBAR

	Print a chart
	Add a chart


An Individual image can be removed from the layout by right-clicking on the image and selecting the “Remove from Chart” option in the pop-up menu.




Multiple images may be dragged from the filmstrip to a single position on any given chart. To reveal multiple images in a given position, right click on the thumbnail image. The following box should appear:



When it says “Send Images”, the plural of Image means there is more than one image. By clicking on “Send Images”, the Transfer Images Window should appear.



Clicking above any image in the filmstrip will show this dialogue which allows one to create a filmstrip window that can float above all open applications. This is handy for charting, etc.



The following information applies only when one is in the “View Exam” Tab:


Double clicking a radiograph that is on the chart will send all radiographs (taken on that chart location) to the “Diagnosis View” Tab in History Mode. Images from all dates taken in that chart position are displayed in the filmstrip. History Mode films is indicated by a blue background on the filmstrip in the Diagnosis View Tab.

This applies to all studies taken on a given patient.

TRANSFER IMAGES WINDOW



This window appears when the user right-clicks on an image and then the “Send Image” or “Send Images” option is clicked.

It provides for copying, previewing, printing, e-mailing, and exporting in a variety of file formats as well as other options.



TRANSFER IMAGES WINDOW


In order to transfer images from tab to tab, between patients, into other file formats, across the internet and to other locations, the Image Transfer System was created. To send an image to a different location, right-click on the image or thumbnail, and select the “Send Image(s)” item from the pop-up menu.



This will open the “Transfer Images” window.

The selected image(s) will be shown as well as the options for where that/those image(s) can be sent. Click on the location to which you wish to send the image(s) and it/they will be sent directly as in the case of Copy to system clipboard, Preview to preview window, or Diagnosis View where it lands directly into and launches the Diagnosis View Tab.

Other options launch further windows from which to work. Send Email launches the available e-mail client (configure email in Options). At the same time, all the selected images are exported as .JPG files, placed in a .ZIP archive and put on the computer desktop. File Export opens a “Save” dialog box with options for multiple file types. Print Image sends the image to the printer.



Save as.... will open a dialog box that allows you to move the image(s) to an alternate patient under a new study. No demographic data will be changed in this instance.

Edit DICOM Header Tags will allow you to change metadata embedded within the image, but it is strongly recommended that this be done only by those experienced with those features.

The EVAsoft Claims button will create a folder under the patient’s name, which in turn will be located in the “Claims” folder. The location of the claims folder is determined by the “EVAsoft Claims” preferences window under Tools > Options. The selected image(s) will then be deposited in the folder designated by the patient’s name. In addition, the file type may be JPEG, TIFF, PNG, or DICOM, as determined in the “EVAsoft Claims” preferences window.

EVAsoft FILE MENU

File

Import Image...
Import DICOM Dir...

Switch Users...
Change Password

Exit

IMPORT IMAGE - Can be used to import an image file.
IMPORT DICOM DIR - Can be used to import DICOM directory patient information and imagefiles.

SWITCH USERS - Allows the user to change the user account that is logged in.

CHANGE PASSWORD - Can only be used by the administrator user to change user account passwords.

EXIT - Will close the software.

EVAsoft EDIT MENU

Edit

Undo
Redo

Tracability Logs
User Accounts
Dentist List
Facility List

Layout Editor
Image Recycle Bin

Custom Overlays
Custom Enhancements

UNDO/REDO - Will undo or redo the last changes made.

TRACEABILITY LOGS - Brings up a window that facilitates tracking activity within the program such as log-ins and outs, new image acquisitions, etc.

**USER ACCOUNTS
DENTIST LIST AND
FACILITY LIST** } Allows the user to edit the information and add new items. Tool is useful for creating new

LAYOUT EDITOR - Acquire Template layouts. Allows the user to undelete images

IMAGE RECYCLE BIN - or completely remove images from the EVAsoft database.

**CUSTOM OVERLAYS
CUSTOM ENHANCEMENTS** } Overlays and Enhancements that have been created by a user can be edited.

: Tools Preview Help

Configuration Tool ...

Update Thumbnails

Sync Images Now

Replicate All Files Locally

Change Language

DICOM Settings

Options

Configuration Tool: Location of device-specific settings

Update Thumbnails: Refreshes the thumbnail images.

Sync Images Now: Prompts the Host to fetch all images from this workstation now.

Replicate All Files Locally: Is used for Network Mode. It will create copies of all image files locally to improve the image loading time.

Change Language: Can be used to toggle between English and other languages.

DICOM Settings: Allows the user to edit the information needed to configurate Send, Receive, and Worklist functions for multiple servers.

Options: This tool contains a number of general settings that the user can choose from. See Appendix D.

SEND Settings -

DICOM Configuration Settings

Send Settings | Worklist Settings | Receive Settings

DICOM Server List

Delete Selected | Ping | Validate

Server Information

DICOM Auto-Send Enable: ☐

Name for Server:

Remote AE Title:

DICOM Server Address:

DICOM Server Port:

Local AE Title: IMAGEWORKS

Update Selected | Add as New Entry

Global Send Settings

Retry Wait Time (in milliseconds): 5000

Done | Cancel

Worklist Settings -

DICOM Configuration Settings

Send Settings | Worklist Settings | Receive Settings

All fields are Required.

Remote AE Title:

Local AE Title: IMAGEWORKS

DICOM Server Address:

DICOM Server Port:

Date: 20150309

Wait Time (s): 10

Write Time (s): 30

Modality: ALL

☐ Show a confirmation window for each worklist patient

Validate

Done | Cancel

Receive Settings -

DICOM Configuration Settings

Send Settings | Worklist Settings | Receive Settings

Changes to these setting will take effect the next time you start the Software.

Folder Name: dcmrcv


Port: 4006

AE Title: IMAGEWORKS

Launch at Start-up: ☐

Done | Cancel


EVAsoft PREVIEW MENU



MANAGE - Allows the user to create or delete preview windows. As the windows are created, they will be appended to the preview menu.

CLEAR ALL WINDOWS - Will clear all images from the active preview windows.

HELP MENU



ONLINE DOCUMENTS AND VIDEO

VIDEO - Will launch the default internet browser and open the www.ImageWorksCorporation.com website and the EVAsoft support page. Requires internet connection.

USER'S MANUAL - Launches the User Manual.

REMOTE TECHNICAL SUPPORT - Opens the remote support login page manned by ImageWorks technical support representatives during normal business hours. Requires internet connection.

LICENSE REGISTRATION - Allows the user to enter the serial number that came with the EVAsoft CD. Based on this number, the software generates a Challenge Code. Both numbers must be supplied to ImageWorks, either online or by phone, to receive the activation code that will register the software installation.

ABOUT EVASOFT - Tells the user information about the software.

ENHANCEMENT TOOL DESCRIPTION

Angle Measurement
 Brightness/Contrast/Gamma
 Crop
 Despeckle Filter
 Edge Enhance
 Equalization
 Flip/Rotate
 Gamma
 Invert
 Linear Stretch
 Local Linear Stretch
 Image Calibration
 Image Measurement
 Sharpen
 Soften
 USM
 Window/Level/Gamma

ANGLE MEASUREMENT - Measures a single angle on the image.
BRIGHTNESS/CONTRAST/GAMMA - This is a combination tool that allows the user to adjust the contrast and brightness without saturating the image. The gamma level can also be adjusted in this tool.

CROP - Select an area of the image to which to apply the image manipulation tools.

DESPECKLE FILTER - Smooth areas in which noise is noticeable while leaving complex areas of the image untouched. The effect is that grain or other "Salt & Pepper" noise is reduced without severely affecting edges.

EDGE ENHANCE - Enhances the local discontinuities at the boundaries of different objects (edges) in the image to make the edges more visible to the human eye.

EQUALIZATION - Linearizes the number of pixels in an image based on the specified (RGB, YUV or GRAY) color space. This can be used to bring out the detail in dark areas of an image.

FLIP/ROTATE - Change the orientation of the image. Options include mirroring horizontally and vertically, as well as rotating by 90, 180 or 270 degrees.

GAMMA - Increase or decrease the level of Gamma correction, which changes the contrast of the mid-range image data. This helps distinguish bone from soft tissue.


INVERT - Inverts the colors in the image so that a positive image becomes a negative and a negative image becomes a positive.

LINEAR STRETCH - Maximize the contrast of the image by stretching a selected portion of the image histogram across all possible values in a linear fashion.

LOCAL LINEAR STRETCH - A very sophisticated linear stretch performed on multiple smaller overlapping areas of the image.

IMAGE CALIBRATION - Enter the known distance, in millimeters, between two points on the image. This is a relative calibration which is not traceable.

IMAGE MEASUREMENT - Measures the distance between up to 10 points on a calibrated image. Due to the possibility of geometric distortion in the radiograph, use caution when interpreting measurements.




ENHANCEMENT TOOL DESCRIPTION (CONT'D)



- SHARPEN - Performs a High-Pass filter operation.
- SOFTEN - Performs a Low-Pass filter operation.
- USM - Unsharpen Mask - A technique used to enhance the edges of an image. The unsharpen mask tends to show more details in areas of an image that are normally more difficult to see.

WINDOW/LEVEL/GAMMA - Adjusts window, level and gamma settings in one user interface.



SOME FILTER ANALOGIES:

LOW PASS -	HIGH PASS -
<ul style="list-style-type: none">• bass• opaque• muted/dull• soft• blurry• pliable	<ul style="list-style-type: none">• treble• transparent• intense/vivid• harsh/edgy• sharp• brittle

ENHANCEMENT TOOL DESCRIPTION (CONT'D)




PSEUDO COLOR AND THRESHOLD - May be found in the "All Tools Window", accessed by clicking on this icon found in the Diagnosis View Tab.



PSEUDO COLOR - Apply color to a grayscale image, with matching grayscale intensities assigned one color. This filter applies only in a temporary manner and cannot be saved.

HISTOGRAM - This shows a graphical representation of the image data.




THRESHOLD - Selects a pixel intensity value to highlight in the image. The user can also select different colors and how far from the original pixel value to highlight. This filter applies only in a temporary manner and cannot be saved.

APPENDIX A

LAYOUT EDITOR

Select Layout Editor from the Edit menu. The following window will open:



To create a custom layout, click on the New Layout button.



If there is an image file you wish to use as the background to the layout, select File > Import Image and import the image into EVAsoft, to any tab. Then in the layout editor, click on the Schematic button and select the imported image file.

Now add Regions. Each region must have a unique name.


Move the Region marker to the corresponding region in the schematic.

APPENDIX B

PRINT PDF REPORTS



From the Patient Tab, clicking on the PDF report button will bring up a small word processor as shown:




APPENDIX B (CONT'D)

PRINT PDF REPORTS

There is a filmstrip at the bottom of this window that contains the radiographs of the current patient/study. Patient, Provider and Facility Information tabs allow this information to be entered easily.

The File drop-down menu provides for opening existing reports, creating new reports, saving as an EVAsoft file, printing a PDF file, or closing the window. Font and Color drop-down menus are for font sizing and color.



APPENDIX C

NETWORK SETUP

EVAsoft can be set up to run in two (2) modes: Stand-Alone OR Network.

The mode of operation should be decided at the time of installation.


Stand-Alone mode Used for any installation that will only use one computer to acquire and view images using EVAsoft. Simply select the Stand-Alone mode during the InstallShield window as shown below. No additional steps are required.


Network mode Used for installations where several different computers will be used to acquire and view images with EVAsoft.

To have multiple computers connect to ONE Shared EVAsoft database, one installation must HOST the database. The Host computer must ALWAYS be on and EVAsoft must ALWAYS be running. The only registration required in network mode is the host. This should be done first before installing any of the clients.

Step One: The first installation should be done on the computer (preferably a server) that will be hosting the Database. Choose the “Network Mode and Host the database” option during the InstallShield window as shown below.

Step Two: All additional computers should be installed as clients, to do that select “Network Mode and connect to already hosted database” option during the InstallShield window as shown below.


 **NOTE:** The mode of operation, whether it is Network or Stand-Alone has nothing to do with EVAsoft’s ability to send DICOM images over an existing network. Network mode is simply to ensure that multiple installations of EVAsoft can utilize a central or common HOST database.



APPENDIX C (CONT'D)

In order for all the CLIENT computers to find the HOST, the user may need to enter the computer name of the Host computer/server.


To find the computer name, right-click on My Computer and select Properties from the pop-up menu. Click on the Computer Name tab, as shown below. The full computer name should be entered on the client computers, if EVAsoft cannot automatically find the Host.




APPENDIX D

TOOLS > OPTIONS:

EVAsoft has a set of option preferences that can be set for various default values used within the software. These can range from pixel sizing on the monitor used, to acquisition preferences, to login and password information, among many other features. Here are the options offered:



The General preferences option allows you to set the number of undo/redo operations. It also will let you choose whether or not a login is required and if so, whether or not a password is required.



The Diagnosis View window allows one to set the number of millimeters per screen pixel. The choice can be made as to whether or not a custom overlay (demographics) will be displayed on acquired images as a default. If so, the type of overlay can be specified here as well.

In addition, there are radio buttons at the bottom that allows one to choose the grid layout. The grid layout can be set to display one, two or four images at a time.


APPENDIX D (CONT'D)

TOOLS > OPTIONS:



The Practice Management Bridge window allows one to choose whether or not the program will be operating in Practice Management Mode as well as whether the program starts in the Patient Tab or Acquire Tab.

The Acquire Tab Options window allows one to choose between FDI or US tooth numbering. The default layout as well as default age for the pedo toothchart and enabling of systems sounds along with desired sound type may be chosen. The traffic Light can also be enabled.




The Acquire Tab Options window allows one to choose between FDI or US tooth numbering. The default layout as well as default age for the pedo toothchart and enabling of systems sounds along with desired sound type may be chosen. The traffic light can also be enabled.


There is also a check box that can be used to automatically restart Twain devices. This is most useful in eliminating the need to constantly click the start button with devices such as cameras.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:




The Default Facility window allows one to choose a facility from the drop-down list.




The Default Dentist window allows one to choose a dentist from the drop-down list.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:




These options should not be altered except by authorized technical support or IT personnel.




The Distributed File System window allows one to choose the location of the volume and its maximum capacity.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:




The EVAsoft Claims window allows one to determine the default path where a folder called "Claims" is located. Whenever an image is sent to EVAsoft Claims, a new folder is generated under the "Claims" folder that is the patient's name. A file type of JPEG, TIFF, PNG or DICOM may also be chosen. This helps facilitate insurance claims.




The Demographics Display window sets size position and spacing of the demographic text overlay on the Image on the screen.

APPENDIX D (CONT'D)


TOOLS > OPTIONS:



The DICOM Export Settings window allows one to choose between the .dic and .dcm extensions OR no file extension at all when exporting an image in the DICOM format. Default is .dcm.



The Patient Setup window allows one to choose whether or not there is automatic incrementing of Patient ID's and what number they start with.


 Note: This is for new patients only.

A drop down menu allows you to set the default patients displayed in the patient tab.
Options are:

- Recent Patients
- Show All Patients
- Patients Over Last Month


APPENDIX D (CONT'D)

TOOLS > OPTIONS:



View Exam options window contains a checkbox that will turn on or off filters that may have been applied from the Diagnosis View Tab. A second check box allows image thumbnails to be displayed in the chart positions.


Having these boxes unchecked allows the charts to load faster.



EVA Preferences in the Options menu is for technical support only. This is only used for EVA Classic sensors.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:



Automated Data Back-Up Utility

EvaSoft offers a process to set automatic backup of your patient information and images. The frequency of back ups as well as the location to which they are stored should be set by the office. Best practice is to set a schedule for this backup, and identify a location that is included in the broader network data back up process for the office.

To enable this back up procedure, click on Tools menu, and select Options from the drop down menu.

In the Options dialogue box, select Back-Up Utility from the menu along the left. Check the "Back-up Enable" box.

Set time and frequency.

Select Browse Backup Path to identify the location of the back up.

This process will create a copy of the critical patient and image information stored in the location you designate. These files should be included as part of your broader back-up system.


ImageWorks strongly recommends that data back-ups are performed on a regular basis.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:




The EVA Select Controls are for technical support use only.



The Panoura 18S Controls are for technical support use only.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:



Setting up functionality requires entering the information into the Email Preferences Dialog accessed through the EVAsoft menu, Tools, Options, email.

Contact your email provider and record the enter the following information for SMTP (this information may be available online as well):

Your email address

Your email Provider's Account Host

Your Email Provider's Account Port


If your email using SSL (yes/no)

Email password authentication



User name (typically just the email address).

Click the Verify button to check your connection to the email server. Click OK to save the settings and return to EVAsoft.

In the filmstrip select one or more images



Hover over image and right mouse click. Then click Send Mail button.



Add the destination email address(s) and modify the body of the email text as required.

APPENDIX D (CONT'D)

TOOLS > OPTIONS:

Options

Outgoing Email Preferences

Outgoing Email Preferences

Account Description Default

Account Email Address

Account Host

Account Port 465

☒ Use Secure Sockets Layer (SSL)

Password Authentication Username

Account Password

Email account status: Not Verified

Verify

OK Cancel

Enter the email settings for your account as supplied by your provider (such as gmail).

Click the Verify button to connect to the email account and check the settings.

EVA SELECT & EVAsoft APPENDIX E

Supplemental Information



Caution: US Federal Law restricts this device to sale by or on the order of a licensed medical professional. Other jurisdictions may have similar laws. Ensure that you are in compliance with the legal requirements for medical devices in your country and locality.

Indications: The EVA SELECT Digital Dental Imaging System is intended to be used with standard dental X-ray systems and computer stations for system operation, archive data storage, digital dental image input capture and enhancement, and patient data and support.

1. COMPUTER REQUIREMENTS:

See website www.imageworkscorporation.com for detailed Computer Requirements.

IMPORTANT: If the package or the contents were damaged in shipping do not use. Document the damage and contacted the shipper immediately.

2. INSTALLATION:

See installation instructions supplied with the sensor.

3. CARE INSTRUCTIONS:

3.1 General Precautions:

The EVA SELECT dental intra oral X-ray sensor is a sophisticated electronic product incorporating advanced technologies. As such, it is to be handled with a high degree of care. Moreover, by design, the EVA SELECT dental intra oral X-ray sensor cannot be used in certain conditions:



Note: This device cannot be used in anesthetic gas environment.

3.2 Cleaning instructions:

Ensure that ALL personnel who use or handle the sensor carefully read and understand the cleaning and disinfection instructions below. Following these recommendations will allow you to clean the sensor safely for both the sensor and your patients and your staff.

3.2.1 Recommended practices:

The EVA SELECT dental intra-oral X-ray sensor is NOT supplied as a sterile medical device. Disinfect the

- sensor before first use and subsequently before using with every new patient. Always use protecting gloves when using, cleaning or disinfecting the sensor.
- Use a new disposable protection barrier at every sensor usage. This protection must be biocompatible
- following the standard ISO 10993-1 or equivalent.

- Cleaning by wiping the sensor's head and the first 5 inches (10 cm) of the sensor cable with a compress moistened with a disinfecting solution is permitted.
- Disinfection by immersion in a disinfecting solution is allowed for the EVA Select.
- Do not immerse the sensor if there are nicks or deep scratches on the sensor head. Only immerse the sensor head and cable, NOT the gray plastic connector section.

3.2.2 Warnings:

- ***The EVA SELECT dental intra-oral X-ray sensor is not designed for high temperatures. DO NOT sterilize the sensor using an autoclave or a UV oven.***
- The sensor housing and the cable are made out plastic materials. Bleach or alcohol content solutions will damage it. Please refer to the recommended disinfecting products list below.
- The sensor head and the cable are watertight, but the Gray Connector Section is NOT. ***DO NOT*** immerse the connector section in the disinfectant for cleaning
- The sensor housing is fragile. ***DO NOT*** clean using inappropriate instruments

3.2.3 Recommended disinfecting solutions:

Do not use aggressive products that may degrade the sensor.

- | | |
|----------------------------------|--|
| • Preferred disinfectants: | ANIOXY TWIN™ (ANIOS Laboratories) PHAGOCID D™ (PHAGOGENE DEC. Laboratories) |
| • Other authorized disinfectants | CIDEX OPA™ (JOHNSON & JOHNSON)
DENTASEPT ultra™ (ANIOS Laboratories)
RELYON PERASAFE™ (Phagogene DEC. Labs)
CAVIWIPES, CAVIWIPES XL (Metrix Research) |
| • Forbidden products: | ALCOHOLS (Isopropyl Alcohol, Methanol) SEKUSID-N™ (ECOLAB PARAGERM Laboratories) SEKUSEPT Easy™ or Aktiv™ (Ecolab Paragerm Labs) FD333™ or FD322™ (DÜRR DENTAL Laboratories) |

3.3 Instructions for use (handling):

Please ensure that ALL personnel who use or handle and work around the sensor carefully read and understand the instructions below. Following these recommendations will ensure you handle the sensor safely and will ensure you do not shorten the sensor's lifespan.

3.3.1 Recommended practices:

Handling:

- Always handle and manipulate the sensor with utmost care.
- Plug and unplug the sensor holding the gray plastic connector (not the cable).
- Ensure that the sensor head never strikes a hard surface (for example falls to the floor from a counter).
- NEVER pull the cable, carefully lift the sensor head.
- Ensure that the cable is not tangled when using it.
DO NOT bend the cable too severely, this can cause permanent damage to the sensor.

Disposables:

- Use the holders (Aiming Aid), protection barriers and other disposables supplied with the sensor or recommended by your dealer.
- Always use the sensor with a holder and protection cap of the proper size.
- To remove the sensor from the holder, grip the sensor carefully and withdraw holder (do not pull out by the cable).

3.3.2 Warnings:

While in use, most accidental damage to the sensor occurs when the cable is lying on the floor or is run across a hallway or other passageway.

Configuration:

The installation should be configured in order to avoid:

- Cables lying on the floor when in use.
- Running cables across a hallway or other trafficked area.

Handling:

The following handling errors could damage your sensor permanently:

- Pinch sensor or cable severely (ex: cable pinched in a drawer)
- Pull or bend cable severely
- Unplug the sensor by pulling the cable
- Drop sensor (see above)
- Rolling over the cable with a chair or walking over it (see above)
- Bite sensor or cable

The following handling errors could be harmful to the operator and/or the patient:

- Connecting or disconnecting the sensor while electronic interface is powered up.
- Touching the connector pins.
- Liquid on the connector (for example for cleaning).
- Using a sensor which is nicked, cracked, deeply scratched on the sensor's head or on the cable or otherwise obviously damaged.

Disposables:

The following handling errors could damage your sensor permanently:

- Removing the protection barrier by pulling the cable.
- Coiling the cable around the holder during use.

3.4 Instructions for storage:

Please carefully read and follow the instructions below to ensure you handle the sensor safely and do not shorten the sensor's lifespan.

3.4.1 Recommended practices:

Sensor storage between two exposures:

- Leave the sensor plugged into the PC
- Store the sensor in a holder (or a hook) fixed on the wall
- Coil the cable in large loops (one or two loops) in order to avoid the cable lying on the floor

Sensor storage between two patients (or for the night):

- If possible, leave the sensor plugged into the PC (same as above)

- If you have to move the sensor to another room:
 - o Unplug the sensor carefully.
 - o Coil the cable in large loops in order to avoid the cable lying on the floor.
 - o Store the sensor in a holder or a hook fixed on the wall
- If you have to store the sensor on a shelf or drawer, package the sensor well to provide cushioning and avoid shifting and make large loops with the cable (nine inches or 20cm diameter if possible).

3.4.2 Warnings:

When not in use most accidental damage to the sensor occurs when the sensor falls off a counter or the cable is too tightly coiled.

Please make sure to avoid the following configurations:

- Using a very tight box to store the sensor.
- Storing the sensor freely on a table, countertop or a shelf.
- Running, or looping the cable near or over drawers (you could pinch the cable).
- Storing several sensors on the same hook.
- Curling the cable in tight loops (for example around gray plastic connector).
- Running over the cable with chair wheels.
- Catching a loop of the cable in a chair leg.

4. DISPOSAL INSTRUCTIONS:

The sensor head contains potentially hazardous substances (material of the scintillator). It must to be treated specifically at disposal. The cable and connector have to be recycled as electronic waste.

5. SAFETY INSTRUCTIONS:

5.1 General safety recommendations:

Carefully read, understand and follow the instructions and warnings in section 3 of this document.

5.2 Specific safety recommendations:

- a) The computer and any associated equipment (such as a USB hub) shall be placed outside the patient's environment - specifically it must be placed more than 5 feet (1.5 meters) away from the patient chair. The operator shall not touch or hold the patient and such other devices at the same time.
- b) The computer and any other associated equipment shall be compliant with IEC 60950 or IEC 60601.
- c) The maximum temperature of the applied part (The total surface of the sensor and the first 5 feet (1.5meters) of the cable are considered as BF 'applied part'.) is below 48 degrees Celsius (118° F). The applied part shall be in contact with patient less than 10 minutes.

5.3 Markings:



BF applied part



Warning symbol: The accompanying documentation may contain warnings and instructions regarding the safety of the device.

IP67

IP67 symbol

The casing of the sensor is IP67. This means that sensor is :

- totally protected against dust,
- protected against the effect of immersion between 15cm and 1m.

5.4 EMC:

The EVA SELECT sensor is considered as non-life-supporting equipment.

While using EVA SELECT sensors adjacent to other equipment, attention must be given to ensure that electromagnetic interference (EMI) does not degrade the performance. In particular, mobile RF communications equipment can affect medical electrical equipment. Please refer to EMC table below.

Usage limitation: The EVA SELECT sensor shall be used with an IEC 60950 or IEC 60601 compliant computer. Any other devices connected between the EVA SELECT sensor and the computer (for example a USB Hub) shall also be compliant with IEC 60950 or IEC 60601. Failure to conform to this may result in degraded electromagnetic compatibility.

5.5 FCC:

Compliant with FCC part 15 class B rules

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.



NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct interference by one or more of the following measures:

- *Reorient or relocate the receiving antenna.*
- *Increase the separation between the equipment and receiver.*

- Connect the equipment into an outlet on circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

5.5 FCC:

Table according to IEC60601-1-2 Ed2 consists of:

The EVA SELECT sensor is suitable for use in the specified electromagnetic environment. The purchaser or user should assure that it is used in an electromagnetic environment as described below:

Emmissions Test	Compliance	Electromagnetic Environment
Radio-Frequency Emissions CISPR11	Group 1	The EVA SELECT sensor does not use radio frequency energy. Therefore, radio frequency emissions are very low and will not interfere with other nearby electronic products. The EVA SELECT sensor is suitable for use in all establishments including domestic establishments and those directly connected to the public low-voltage power supply networks that supply buildings used for domestic purposes.
Radio-Frequency Emissions CISPR11	Class B	
Harmonic Emissions IEC 61000-3-2	Not Applicable	
Voltage fluctuation and flicker test IEC 61000-3-3	Not Applicable	

The EVA SELECT sensor is suitable for use in the specified electromagnetic environment. The purchaser or user should assure that it is used in an electromagnetic environment as described below:

Immunity Test	IEC 60601-1-2 test level	Compliance test level	Electromagnetic Environment
Electronic discharge (ESD) IEC 61000-4-2	+/-6kV contact discharges +/-8kV air discharges	6kV contact discharges 8kV air discharges	Floors are wood, concrete, or ceramic tile, or floors are covered with synthetic material and the relative humidity is at least 30 percent.
Electronic fast transient IEC61000-4-4	+/-2kV : power supply lines +/- 1kV : input/output lines	+/-2kV : power supply lines +/- 1kV : input/output lines	Mains power quality is that of a typical commercial and/or hospital environment.
Power frequency (50/60Hz) magnetic field IEC61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields are at level characteristics of a typical location in a typical commercial and/or hospital environment.
Surge IEC61000-4-5	+/-1kV : differential mode +/- 2kV : common mode	Not Applicable	

Voltage dips and short interruption. IEC61000-4-11	>95% dip during 0.5cycle 60% dip during 5 cycles 30% dip during 25 cycles >95% dip during 5s	Not Applicable	
---	---	----------------	--

The EVA SELECT sensor is suitable for use in the specified electromagnetic environment. The purchaser or user should assure that it is used in an electromagnetic environment as described below:

Recommended separation distance for portable and mobile RF communications equipment IEC 60601-4-2


Frequency of Transmitter Equation	150kHz to 80Mhz $d = [3.5 / V1] \sqrt{P}$ Separation distance	80MHz to 800Mhz $d = [3.5 / E1] \sqrt{P}$ Separation distance	800MHz to 2.5GHz $d = [3.5 / E1] \sqrt{P}$ Separation distance
Rated Maximum output power of transmitter (watts)			
0.01	0.12	0.12	0.23
0.1	0.37	0.37	0.74
1	1.17	1.17	2.34
10	3.69	3.69	7.38
100	11.67	11.67	23.34

For transmitters with maximum output powered not in the above table, the recommended separation distance (in meter) can be estimated using the formula at the proper frequency, where P is the maximum power in watts (W), E1 is 3 (V/m) and V1 is 3 (V).

NOTE 1: At 80 MHz and at 800 MHz, the distance of the above frequency range will apply.

NOTE 2: These guidelines may not apply in every situation. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

The EVA SELECT sensor is suitable for use in the specified electromagnetic environment. The purchaser or user should assure that it is used in an electromagnetic environment as described below:

Immunity Test	IEC 60601-1-2 Test level	Compliance Level	Electromagnetic Environment
Conducted RF IEC61000-4-6 Radiated RF IEC61000-4-3	3V 150kHz to 80Mhz 3V/m 80MHz to 2.5Ghz	V1 = 3V E1= 3V/3	<p>Portable and mobile RF communication equipments are used no closer to any part of EVA SELECT sensor, including cables, than the recommended separation distance calculated from the equation appropriate for the frequency of the transmitter. Recommended separation distance $d = [3.5 / V1] \text{ eP}$</p> <p>$d = [3.5 / E1] \text{ eP}$ 80 MHz to 800 MHz $d = [3.5 / E1] \text{ eP}$ 800 MHz to 2.5 GHz</p> <p>Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). File strengths from fixed RF transmitters, as determined by electromagnetic site survey,* are less than the compliance level in each frequency range.**Interference may occur in the vicinity of equipment marked with the following symbol :</p> 

NOTE 1 At 80 MHz and at 800 MHz, the distance of the above frequency range will apply.

NOTE 2 These guidelines may not apply in all the situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

* Field strengths from fixed transmitters, such as base stations for cellular telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast cannot be estimated accurately. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be performed. If the measured field strength exceeds the RF compliance level above, observe EVA SELECT sensors to verify normal operation in each use location. If abnormal performance is observed, additional measures may be necessary, such as reorientation or relocation of the EVA SELECT sensors.

** Over the frequency range 150kHz to 80Mhz, field strength is less than V1 V/m.

6. OPERATING RANGE AND ABSOLUTE MAXIMUM RATING

6.1 Operating range:

Room temperature: + 5 °C to + 35 °C

Room humidity: + 5 % to + 85 %

Operating temperature: + 5 °C to + 40 °C

6.2 Absolute maximum ratings:

Storage temperature: - 10 °C to + 50 °C

Storage humidity: 10 % to 80%

Transport temperature: - 10°C to + 50 °C

Transport humidity: 10 % to 80 %

Transport ambient pressure: 50 kPa to 106 kPa

Maximum applied voltages: 5V (USB)

Maximum total dose irradiation: 20 Gy (over product life)

Perishability: Not applicable (products are not perishable Foodstuffs)

7. TECHNICAL SPECIFICATIONS (subject to change):

Sensor Type: CMOS

Interface Type: USB 2.0

Sensor Sizes:

Size 1 (Pedeo) 36.73 mm (L) x 24.35 mm (W) x 5.41 mm (Thick)

Size 2 (Adult) 42.80 mm (L) x 30.49 mm (W) x 5.41 mm (Thick)

Image Acquisition Area:

Size 1 (Pedeo) 30.02 mm (L) x 19.95 mm (W)

Size 2 (Adult) 36.00 mm (L) x 25.99 mm (W)

Capsule Cable length: 3 meters (9.8feet)

Pixel Size: 19 microns

Sensor Resolution:

26.3 lp/mm (theoretical optical)

20.0 lp/mm (actual)

Sensor Dynamic Range: 12-bit (0 to 4095)



ImageWorks

Generations of Imaging

8 Westchester Plaza, Suite 112, Elmsford, NY
10523

Tel (914) 592-6100

www.ImageWorksCorporation.com